

Sound level Meter Calibration report

Instrument Calibrated

Description : Sound Level Meter
Manufacturer : ACO
Model : 6226
Serial No. : 180139

Environment Condition

Temperature : (24±3)°C
Relative Humidity : (55±15)%
Ambient Pressure : (1008±5)hPa

Standards Used

- Sound Level Calibrator Delta ohm HD 2020 S/N: 15030521
- Digital Multi meter Fluke 289/FVF S/N 67845
- Computer and Software Dell Vostro 5460 S/N 9CWD1Y1

Calibrated Result

Method	Standard Reference (dB)	Reading (dB)	Error (dB)	After Adjustment ± (dB)
Sound Calibrator HD-2020	114.0	113.0	-1.0	114.0

CALIBRATED BY : <i>Santi Meemana</i>	DATE : 17/2/20	 JIRANATE ASSOCIATES CO., LTD
APPROVED BY : <i>Santi Meemana</i>	DATE : 17/2/20	
For more information please contact : Santi Meemana Tel : 086-303-3140		
63/14-15 Petkasem 7 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28680812 Fax:(66) 0-2868-1889		

Sound level Meter Calibration report

Instrument Calibrated

Description : Sound Level Meter
Manufacturer : ACO
Model : 6226
Serial No. : 180076

Environment Condition

Temperature : (24±3)°C
Relative Humidity : (55±15)%
Ambient Pressure : (1008±5)hPa

Standards Used

- Sound Level Calibrator Delta ohm HD 2020 S/N: 15030521
- Digital Multi meter Fluke 289/FVF S/N 67845
- Computer and Software Dell Vostro 5460 S/N 9CWD1Y1

Calibrated Result

Method	Standard Reference (dB)	Reading (dB)	Error (dB)	After Adjustment ± (dB)
Sound Calibrator HD-2020	114.0	114.2	+0.2	114.0

CALIBRATED BY : <i>sumanon</i>	DATE : 17/2/20	 JIRANATE ASSOCIATES CO., LTD
APPROVED BY : <i>Director Office</i>	DATE : 17/2/20	
For more information please contact : Santi Meemana Tel : 086-303-3140		
63/14-15 Petkasem 7 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28680812 Fax:(66) 0-2868-1889		



JIRANATEE ASSOCIATES CO., LTD

บริษัท จิรนิติ แอสโซซิเอตส์ จำกัด (สำนักงานใหญ่)

JIRANATEE ASSOCIATES CO.,LTD (Head office)

63/14-15,67/35-36 ซอยเพชรเกษม7,7/1 ถนนเพชรเกษม แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600 โทร: (66) 02-8680812#13 แฟกซ์: (66) 02-8680860
63/14-15,67/35-36 Soi Petchkasem7,7/1, Petchkasem Rd, Watthapra, Bangkokyai, Bangkok 10600 Thailand Tel : (66) 02-8680812#13 Fax : (66) 02-8680860



Sound level Meter Calibration report

Instrument Calibrated

Description : Sound Level Meter
 Manufacturer : ACO
 Model : 6226
 Serial No. : 180077

Environment Condition

Temperature : (24±3)°C
 Relative Humidity : (55±15)%
 Ambient Pressure : (1008±5)hPa

Standards Used

- Sound Level Calibrator Delta ohm HD 2020 S/N: 15030521
- Digital Multi meter Fluke 289/FVF S/N 67845
- Computer and Software Dell Vostro 5460 S/N 9CWD1Y1

Calibrated Result

Method	Standard Reference (dB)	Reading (dB)	Error (dB)	After Adjustment ± (dB)
Sound Calibrator HD-2020	114.0	114.1	+0.1	114.0

CALIBRATED BY : <i>gamm</i>	DATE : 17/2/20	 JIRANATEE ASSOCIATES CO., LTD.
APPROVED BY : <i>gamm</i>	DATE : 17/2/20	
For more information please contact : Santi Meemana Tel : 086-303-3140		
63/14-15 Petkasem 7 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28680812 Fax:(66) 0-2868-1889		